

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LEB

Date: 07/24/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:34
INTERNAL STANDARD	VERIFIED	18:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:34
BLANK TEST	0.00	18:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:35
BLANK TEST	0.00	18:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:37
BLANK TEST	0.00	18:38

