

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 08/08/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:28
INTERNAL STANDARD	VERIFIED	19:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:28
BLANK TEST	0.00	19:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:29
BLANK TEST	0.00	19:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:31
BLANK TEST	0.00	19:31

