

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 08/15/2024

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:47
INTERNAL STANDARD	VERIFIED	14:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:48
BLANK TEST	0.00	14:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:49
BLANK TEST	0.00	14:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:50
BLANK TEST	0.00	14:51

