

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 08/21/2024

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:29
INTERNAL STANDARD	VERIFIED	23:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:29
BLANK TEST	0.00	23:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:30
BLANK TEST	0.00	23:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:31
BLANK TEST	0.00	23:32

