

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 09/05/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:49
INTERNAL STANDARD	VERIFIED	18:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:49
BLANK TEST	0.00	18:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:50
BLANK TEST	0.00	18:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:52
BLANK TEST	0.00	18:52

