

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 09/20/2024

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:37
INTERNAL STANDARD	VERIFIED	23:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:37
BLANK TEST	0.00	23:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:38
BLANK TEST	0.00	23:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:39
BLANK TEST	0.00	23:40

