

**SUPERVISOR TEST**

-----  
 ONEIDA COUNTY  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 113606  
 -----

Date: 09/25/2024

OPERATOR NAME:  
 STEVEN P EILERS  
 Permit #: 49321  
 Operator Agency/Dept: ONEIDA CO. SO  
 Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:48
INTERNAL STANDARD	VERIFIED	19:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:48
BLANK TEST	0.00	19:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:49
BLANK TEST	0.00	19:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:51
BLANK TEST	0.00	19:51

