

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 10/09/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:47
INTERNAL STANDARD	VERIFIED	00:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:47
BLANK TEST	0.00	00:48
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:49
BLANK TEST	0.00	00:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:50
BLANK TEST	0.00	00:51

