

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 10/23/2024

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:07
INTERNAL STANDARD	VERIFIED	19:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:07
BLANK TEST	0.00	19:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:08
BLANK TEST	0.00	19:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:10
BLANK TEST	0.00	19:10

