

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LEB

Date: 10/30/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:09
INTERNAL STANDARD	VERIFIED	21:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:09
BLANK TEST	0.00	21:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:10
BLANK TEST	0.00	21:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:11
BLANK TEST	0.00	21:12

