

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 11/14/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:55
INTERNAL STANDARD	VERIFIED	23:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:55
BLANK TEST	0.00	23:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:56
BLANK TEST	0.00	23:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:58
BLANK TEST	0.00	23:58

