

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 11/20/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:30
INTERNAL STANDARD	VERIFIED	04:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:30
BLANK TEST	0.00	04:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:31
BLANK TEST	0.00	04:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:33
BLANK TEST	0.00	04:33

