

**SUPERVISOR TEST**

-----  
ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606  
-----

Date: 11/27/2024

OPERATOR NAME:  
JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:23
INTERNAL STANDARD	VERIFIED	19:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:23
BLANK TEST	0.00	19:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:24
BLANK TEST	0.00	19:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:26
BLANK TEST	0.00	19:26

