

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

LEB

Date: 12/04/2024

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:43
INTERNAL STANDARD	VERIFIED	19:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:43
BLANK TEST	0.00	19:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:44
BLANK TEST	0.00	19:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:46
BLANK TEST	0.00	19:46

