

SUPERVISOR TEST

1EB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 12/18/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:02
INTERNAL STANDARD	VERIFIED	05:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:03
BLANK TEST	0.00	05:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:04
BLANK TEST	0.00	05:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:05
BLANK TEST	0.00	05:06

