

SUPERVISOR TEST

LFB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 12/25/2024

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:32
INTERNAL STANDARD	VERIFIED	16:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:33
BLANK TEST	0.00	16:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:34
BLANK TEST	0.00	16:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:35
BLANK TEST	0.00	16:36

