

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

LEB

Date: 01/11/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:19
INTERNAL STANDARD	VERIFIED	03:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:20
BLANK TEST	0.00	03:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:21
BLANK TEST	0.00	03:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:22
BLANK TEST	0.00	03:23

