

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

LEB

Date: 01/17/2024

OPERATOR NAME:  
JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 23350

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 05:14 |
| INTERNAL STANDARD       | VERIFIED | 05:14 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 05:14 |
| BLANK TEST              | 0.00     | 05:15 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 05:15 |
| BLANK TEST              | 0.00     | 05:16 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 05:16 |
| BLANK TEST              | 0.00     | 05:17 |

