

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 08/21/2024

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:23
INTERNAL STANDARD	VERIFIED	23:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:23
BLANK TEST	0.00	23:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:24
BLANK TEST	0.00	23:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:26

RFI Detected

