

SUPERVISOR TEST

ONEIDA CO.
SHERIFF'S DEPT
CONTRACT SIGN dmt: 47906

Date: 08/28/2024

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:17
INTERNAL STANDARD	VERIFIED	20:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	20:18
BLANK TEST	0.00	20:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:19
BLANK TEST	0.00	20:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	20:20
BLANK TEST	0.00	20:21

