

SUPERVISOR TEST

LE13

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 06/12/2024

OPERATOR NAME:  
JASON L OSIKA  
Permit #: 35567  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:23
INTERNAL STANDARD	VERIFIED	22:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:23
BLANK TEST	0.00	22:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:24
BLANK TEST	0.00	22:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:26
BLANK TEST	0.00	22:27

