

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 09/03/2025

OPERATOR NAME:
BRADLY P YATES
Permit #: 53896
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:11
INTERNAL STANDARD	VERIFIED	04:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:11
BLANK TEST	0.00	04:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:13
BLANK TEST	0.00	04:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:14
BLANK TEST	0.00	04:15

LEB
Solution
Change

