

# SOLUTION CHANGE

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 04/01/2025

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 25040

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 18:10 |
| INTERNAL STANDARD       | VERIFIED | 18:10 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 18:10 |
| BLANK TEST              | 0.00     | 18:11 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 18:11 |
| BLANK TEST              | 0.00     | 18:12 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 18:12 |
| BLANK TEST              | 0.00     | 18:13 |

