

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 10/01/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:16
INTERNAL STANDARD	VERIFIED	00:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:16
BLANK TEST	0.00	00:17
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:17
BLANK TEST	0.00	00:18
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:18
BLANK TEST	0.00	00:19

