

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 02/01/2025

OPERATOR NAME:
AARON M ALSHAMAN
Permit #: 50807
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:49
INTERNAL STANDARD	VERIFIED	18:49
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:50
BLANK TEST	0.00	18:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:51
BLANK TEST	0.00	18:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:52
BLANK TEST	0.00	18:53

