

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 12/02/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:35
INTERNAL STANDARD	VERIFIED	07:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	07:36
BLANK TEST	0.00	07:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	07:37
BLANK TEST	0.00	07:38
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	07:38
BLANK TEST	0.00	07:39

LEB Solution
change

