

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 03/01/2025

OPERATOR NAME:
AARON M ALSHAMAN
Permit #: 50807
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:05
INTERNAL STANDARD	VERIFIED	22:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:06
BLANK TEST	0.00	22:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:07
BLANK TEST	0.00	22:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:08
BLANK TEST	0.00	22:09

