

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 08/01/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:14
INTERNAL STANDARD	VERIFIED	18:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:15
BLANK TEST	0.00	18:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:16
BLANK TEST	0.00	18:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:17
BLANK TEST	0.00	18:18

