

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 05/02/2025

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:56
INTERNAL STANDARD	VERIFIED	15:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:56
BLANK TEST	0.00	15:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:58
BLANK TEST	0.00	15:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:59
BLANK TEST	0.00	16:00

LEB

