

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 08/27/2025

OPERATOR NAME:
BRADLY P YATES
Permit #: 53896
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:55
INTERNAL STANDARD	VERIFIED	21:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:55
BLANK TEST	0.00	21:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:57
BLANK TEST	0.00	21:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:58
BLANK TEST	0.00	21:59

