

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 12/10/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:36
INTERNAL STANDARD	VERIFIED	19:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:36
BLANK TEST	0.00	19:37
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:37
BLANK TEST	0.00	19:38
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:39
BLANK TEST	0.00	19:40

