

SUPERVISOR TEST




ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 09/16/2025

OPERATOR NAME:
BRADY A WOOD
Permit #: 52004
Operator Agency/Dept: ONEIDA COUNTY SO
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:58
INTERNAL STANDARD	VERIFIED	23:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:59
BLANK TEST	0.00	00:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:00
BLANK TEST	0.00	00:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:01
BLANK TEST	0.00	00:02



9W?} ♠♠ ♡@&" "♦↑!Ç@B⊙⊙⊙<>!≡