

# SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

LEB

Date: 09/24/2025

## OPERATOR NAME:

BRADY A WOOD

Permit #: 52004

Operator Agency/Dept: ONEIDA CO. SO

Reference Standard #: 25170

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 18:53 |
| INTERNAL STANDARD       | VERIFIED | 18:54 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.09     | 18:54 |
| BLANK TEST              | 0.00     | 18:55 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.09     | 18:55 |
| BLANK TEST              | 0.00     | 18:56 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.09     | 18:56 |
| BLANK TEST              | 0.00     | 18:57 |

