

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 10/22/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:44
INTERNAL STANDARD	VERIFIED	03:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:44
BLANK TEST	0.00	03:45
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:46
BLANK TEST	0.00	03:46
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:47
BLANK TEST	0.00	03:48

