

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 11/06/2025

OPERATOR NAME:
JAYME CONNOLLY
Permit #: 53524
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:56
INTERNAL STANDARD	VERIFIED	14:56
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	14:57
BLANK TEST	0.00	14:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:58
BLANK TEST	0.00	14:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	14:59
BLANK TEST	0.00	15:00

LEB

