

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 03/19/2025

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:40
INTERNAL STANDARD	VERIFIED	22:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:41
BLANK TEST	0.00	22:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:42
BLANK TEST	0.00	22:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	22:43
BLANK TEST	0.00	22:44

