

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 01/29/2025

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:44
INTERNAL STANDARD	VERIFIED	19:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:45
BLANK TEST	0.00	19:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:46
BLANK TEST	0.00	19:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:47
BLANK TEST	0.00	19:48

