

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 113606

Date: 02/05/2025

OPERATOR NAME:  
AARON M ALSHAMAN  
Permit #: 50807  
Operator Agency/Dept: ONEIDA CO. SO  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:22
INTERNAL STANDARD	VERIFIED	18:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:22
BLANK TEST	0.00	18:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:24
BLANK TEST	0.00	18:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:25
BLANK TEST	0.00	18:26

