

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 06/06/2025

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:23
INTERNAL STANDARD	VERIFIED	17:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:23
BLANK TEST	0.00	17:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:24
BLANK TEST	0.00	17:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:25
BLANK TEST	0.00	17:26

