

SUPERVISOR TEST

LE13

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 01/08/2025

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:51
INTERNAL STANDARD	VERIFIED	20:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:51
BLANK TEST	0.00	20:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:52
BLANK TEST	0.00	20:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	20:53
BLANK TEST	0.00	20:54

