

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 10/08/2025

OPERATOR NAME:
BRADLY P YATES
Permit #: 53896
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:32
INTERNAL STANDARD	VERIFIED	22:32
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:32
BLANK TEST	0.00	22:33
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:33
BLANK TEST	0.00	22:34
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:35
BLANK TEST	0.00	22:35

LEB weekly

