

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 12/17/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:42
INTERNAL STANDARD	VERIFIED	06:42
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:42
BLANK TEST	0.00	06:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:43
BLANK TEST	0.00	06:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:44
BLANK TEST	0.00	06:45

