

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 10/15/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:27
INTERNAL STANDARD	VERIFIED	18:28
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:28
BLANK TEST	0.00	18:29
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:29
BLANK TEST	0.00	18:30
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:30
BLANK TEST	0.00	18:31

