

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 11/15/2025

OPERATOR NAME:
ELDAR MUJIC
Permit #: 54489
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:12
INTERNAL STANDARD	VERIFIED	18:13
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:13
BLANK TEST	0.00	18:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:14
BLANK TEST	0.00	18:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:15
BLANK TEST	0.00	18:16

