

SUPERVISOR TEST

LEB

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 113606

Date: 01/23/2025

OPERATOR NAME:
 JASON L OSIKA
 Permit #: 35567
 Operator Agency/Dept: ONEIDA CO. SO
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:37
INTERNAL STANDARD	VERIFIED	18:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:37
BLANK TEST	0.00	18:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:38
BLANK TEST	0.00	18:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:40
BLANK TEST	0.00	18:41

