

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 02/19/2025

OPERATOR NAME:
AARON M ALSHAMAN
Permit #: 50807
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:44
INTERNAL STANDARD	VERIFIED	23:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:44
BLANK TEST	0.00	23:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:45
BLANK TEST	0.00	23:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	23:46
BLANK TEST	0.00	23:47

