

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 02/26/2025

LEB

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:38
INTERNAL STANDARD	VERIFIED	19:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:39
BLANK TEST	0.00	19:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:40
BLANK TEST	0.00	19:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:41
BLANK TEST	0.00	19:42

