

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 03/12/2025

LEB

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:00
INTERNAL STANDARD	VERIFIED	19:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:00
BLANK TEST	0.00	19:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:01
BLANK TEST	0.00	19:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:02
BLANK TEST	0.00	19:03

