

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 03/05/2025

OPERATOR NAME:
AARON M ALSHAMAN
Permit #: 50807
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:32
INTERNAL STANDARD	VERIFIED	18:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:32
BLANK TEST	0.00	18:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:33
BLANK TEST	0.00	18:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:34
BLANK TEST	0.00	18:35

