

SUPERVISOR TEST

LEB

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 113606

Date: 04/16/2025

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. SO
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:57
INTERNAL STANDARD	VERIFIED	17:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:57
BLANK TEST	0.00	17:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:58
BLANK TEST	0.00	17:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	18:00
BLANK TEST	0.00	18:01

